

CLIENT:
Abecca Healthcare Ltd.
82 Sedcole St.
Pahiatua 4910

SAMPLE RECEIVED FROM:
Abecca Healthcare Ltd.

Date Received: 14.5.25

Attn.: Joseph Beatson

SAMPLE DESCRIPTION:
Abecca®, Armadillo™ mattresses & bed bases with
Limbershield™ coating.

Client Order No.:

Client Reference:

1 of 6

BS 7177:2008+ A1:2011 SPECIFICATION FOR RESISTANCE TO IGNITION OF MATTRESSES, DIVANS AND BED BASES.

Date Tested: 27/5/2025

Conditioned at: 50 ± 5% RH. 23 ± 2°C

Water soaking procedure not carried out

Sample tested without bed covers or pillows

BSEN 597-1:1995 Furniture- Assessment of the ignitability of mattresses and upholstered bed bases.

Smouldering Cigarette Test:

Result:

Non-Ignition

Summary Table - Criteria of Ignition:

Smouldering Criteria	1 st Test	2 nd Test
Unsafe escalation (3.1 a)	No	No
Test assembly consumed (3.1 b)	No	No
Smouldered through full thickness (3.1 c)	No	No
Smouldered more than 1 hour (3.2 d)	No	No
Charring more than 50 mm from source (3.1 e)	No	No
Flaming criteria		
Occurrence of flames (3.2)	No	No

"THIS REPORT APPLIES ONLY TO THE SAMPLES TESTED"

Samples and their identifying descriptions have been provided by the client unless otherwise stated. NZWTA Ltd makes no warranty, implied or otherwise as to the source of the tested samples. The above results are designed to provide THE CLIENT WITH GUIDANCE INFORMATION ONLY. This document shall not be reproduced except in full.



C Judan
Key Technical Person

28/05/2025

CLIENT:
Abecca Healthcare Ltd.
82 Sedcole St.
Pahiatua 4910

SAMPLE RECEIVED FROM:
Abecca Healthcare Ltd.

Date Received: 14.5.25

Attn.: Joseph Beatson

SAMPLE DESCRIPTION:
Abecca®, Armadillo™ mattresses & bed bases with
Limbershield™ coating.

Client Order No.:

Client Reference:

2 of 6

BSEN 597-2:1995 Furniture- Assessment of the ignitability of mattresses and upholstered bed bases.

Butane Flame Test:

Flame height 35mm
Ignition application time 15 sec


Result: Non-Ignition

Summary Table - Criteria of Ignition:

Smouldering Criteria	1st Test	2nd Test
Unsafe escalation (3.1 a)	No	No
Test assembly consumed (3.1 b)	No	No
Smouldered through full thickness (3.1 c)	No	No
Smouldered more than 1 hour (3.1 d)	No	No
Charring more than 50 mm from source (3.1 e)	No	No
Flaming criteria		
Unsafe escalation (3.2 a)	No	No
Test assembly consumed (3.2 b)	No	No
Burning through the thickness or to extremities (3.2 c)	No	No
Flaming for more than 120 seconds (3.2 d)	No	No

"THIS REPORT APPLIES ONLY TO THE SAMPLES TESTED"

Samples and their identifying descriptions have been provided by the client unless otherwise stated. NZWTA Ltd makes no warranty, implied or otherwise as to the source of the tested samples. The above results are designed to provide THE CLIENT WITH GUIDANCE INFORMATION ONLY. This document shall not be reproduced except in full.



C Judan
Key Technical Person

28/05/2025

CLIENT:
Abecca Healthcare Ltd.
82 Sedcole St.
Pahiatua 4910

SAMPLE RECEIVED FROM:
Abecca Healthcare Ltd.

Date Received: 14.5.25

Attn.: Joseph Beatson

SAMPLE DESCRIPTION:
Abecca®, Armadillo™ mattresses & bed bases with
Limbershield™ coating.

Client Order No.:

Client Reference:

3 of 6

BS 6807:2006 Assessment of the ignitability of mattresses, upholstered divans and upholstered bed bases with flaming types of primary and secondary sources of ignition.

Crib 7 Test: constructed to standard **BS 6807: 2006 Section 9.**

Result: Ignition Above Sample

Non-Ignition

Summary Table - Criteria of Ignition:

Smouldering Criteria	1st Test	2nd Test
Unsafe escalation (4.1 a)	No	No
Assembly consumed or smoulders – i.e. either side or full thickness (4.1 b)	No	No
Smouldering time- 60 minutes or more (4.1 d)	No	No
Discoloured char more than 100mm from source position- top only (4.1 e)	No	No
Flaming criteria		
Unsafe escalation (4.2 a)	No	No
Assembly consumed (4.2 b)	No	No
Flame front to outside edges or full thickness (4.2 c)	No	No
Ignition 7 – flaming more than 13 minutes after ignition of the crib (4.2 e)	No	No
Ignition 7 - Debris causes an isolated floor fire <10 minutes (4.2 g)	No	No

"THIS REPORT APPLIES ONLY TO THE SAMPLES TESTED"

Samples and their identifying descriptions have been provided by the client unless otherwise stated. NZWTA Ltd makes no warranty, implied or otherwise as to the source of the tested samples. The above results are designed to provide THE CLIENT WITH GUIDANCE INFORMATION ONLY. This document shall not be reproduced except in full.



C Judan
Key Technical Person

28/05/2025

CLIENT:
Abecca Healthcare Ltd.
82 Sedcole St.
Pahiatua 4910

Attn.: Joseph Beatson

Client Order No.:

SAMPLE RECEIVED FROM:
Abecca Healthcare Ltd.

SAMPLE DESCRIPTION:
Abecca®, Armadillo™ mattresses & bed bases with
Limbershield™ coating.

Client Reference:

Date Received: 14.5.25

4 of 6

BS 6807:2006 Assessment of the ignitability of mattresses, upholstered divans and upholstered bed bases with flaming types of primary and secondary sources of ignition.

Crib 7 Test: constructed to standard **BS 6807: 2006 Section 9.**

Result: Ignition from Below Sample


Non-Ignition

Summary Table - Criteria of Ignition:

Smouldering Criteria	1st Test	2nd Test
Unsafe escalation (4.1 a)	No	No
Assembly consumed or smoulders – i.e. either side or full thickness (4.1 b)	No	No
Smouldering time- 60 minutes or more (4.1 d)	No	No
Flaming criteria		
Unsafe escalation (4.2 a)	No	No
Assembly consumed (4.2 b)	No	No
Flame front to outside edges or full thickness (4.2 c)	No	No
Ignition 7 – flaming more than 13 minutes after ignition of the crib (4.2 e)	No	No
Ignition 7 - Debris causes an isolated floor fire <10 minutes (4.2 g)	No	No

“THIS REPORT APPLIES ONLY TO THE SAMPLES TESTED”

Samples and their identifying descriptions have been provided by the client unless otherwise stated. NZWTA Ltd makes no warranty, implied or otherwise as to the source of the tested samples. The above results are designed to provide THE CLIENT WITH GUIDANCE INFORMATION ONLY. This document shall not be reproduced except in full.



C Judan
Key Technical Person

28/05/2025

CLIENT:
Abecca Healthcare Ltd.
82 Sedcole St.
Pahiatua 4910

SAMPLE RECEIVED FROM:
Abecca Healthcare Ltd.

Date Received: 14.5.25

Attn.: Joseph Beatson

SAMPLE DESCRIPTION:
Abecca®, Armadillo™ mattresses & bed bases with
Limbershield™ coating.

Client Order No.:

Client Reference:

5 of 6

BS 7177:2008+ A1:2011 SPECIFICATION FOR RESISTANCE TO IGNITION OF MATTRESSES, DIVANS AND BED BASES.

Continued:

Additional testing on the filling material according to the Furniture and Furnishings (Fire) (Safety) Regulations 1988.

These test methods refer to the earlier British Standards BS 5852-2:1982 and BS 6807: 1986

Annex A: Polyurethane foam in block or cushion form - crib 5

Conditioned at: 50 ± 5% RH. 23 ± 2°C

Water soaking procedure not carried out

Samples tested under 100% flame retardant polyester fabric (mass: 220 g/m² ± 5%)

Result:

Non-Ignition

Summary Table: Criteria of Ignition

Chair configuration as per testing to BS 5852-2:1982 (Annex A - crib 5 / Annex B, C, D - Source 2)

Smouldering Criteria	1 st Test	2 nd Test
Smouldering time (more than 60 minutes?) (4.1 b)	No	No
Unsafe escalation (4.1 c)	No	No
Assembly consumed (4.1 d)	No	No
Smouldering to outside edges or full thickness 4.1 (e)	No	No
Discoloured char more than 100 mm from source position- top only (4.1 f)	No	No
Flaming criteria		
Burning time (more than 10 minutes?) (4.2 b)	No	No
Unsafe escalation (4.2 d)	No	No
Assembly consumed (4.2 e)	No	No
Flame front to outside edges or full thickness (4.2 f)	No	No
Burning time (more than 10 minutes?) (4.2 e)	No	No

This sample meets the requirement for a High Hazard and Very High Hazard Categories.

High Hazard = off-shore installations, hotels, hostels, old peoples' homes, certain hospital wards

Very High Hazard = Prison Cells and locked psychiatric accommodation.

"These test results relate only to the ignitability of the combination under the particular conditions of the test. They are not intended as a means of assessing the full potential fire hazard of the mattress, mattress pad or upholstered bed base in use".

"THIS REPORT APPLIES ONLY TO THE SAMPLES TESTED"

Samples and their identifying descriptions have been provided by the client unless otherwise stated. NZWTA Ltd makes no warranty, implied or otherwise as to the source of the tested samples. The above results are designed to provide THE CLIENT WITH GUIDANCE INFORMATION ONLY. This document shall not be reproduced except in full.



C Judan
Key Technical Person

28/05/2025



All tests reported herein
have been performed in
accordance with the
laboratory's scope of
accreditation

No. 1568310.5

(Please quote this number in all correspondence)

CLIENT:
Abecca Healthcare Ltd.
82 Sedcole St.
Pahiatua 4910

SAMPLE RECEIVED FROM:
Abecca Healthcare Ltd.

Date Received: 14.5.25

Attn.: Joseph Beatson

SAMPLE DESCRIPTION:
Abecca®, Armadillo™ mattresses & bed bases with
Limbershield™ coating.

Client Order No.:

Client Reference:

6 of 6



"THIS REPORT APPLIES ONLY TO THE SAMPLES TESTED"

Samples and their identifying descriptions have been provided by the client unless otherwise stated. NZWTA Ltd makes no warranty, implied or otherwise as to the source of the tested samples. The above results are designed to provide THE CLIENT WITH GUIDANCE INFORMATION ONLY. This document shall not be reproduced except in full.

C Judan
Key Technical Person

28/05/2025